

Amendment No. 1

Date: 18/04/2018

Subject: Amendment to the Tender Enquiry Document.

Ref: Tender no. HLL/HMC/Hospital Furniture/18-19/003 dated 13.04.2018

The following changes have been incorporated in the referred tender.

SECTION-VII TECHNICAL SPECIFICATIONS

Sl. No 01 Hospital Bed

Existing specification:

7. Wheels

a) The bed shall be mounted on four numbers **Thermoplastic, Rubber, Polyurethane with nylon center castors** of 100mm x 32mm & shall be made out of semi round rubber hardness 85/90, Pressed fitted over the virgin polyamide pulling.

Read as:

a) The bed shall be mounted on four numbers Thermoplastic, Rubber, Polyurethane with nylon center castors of 100mm to 125mm x 32mm & shall be made out of semi round rubber hardness 85/90, Pressed fitted over the virgin polyamide pulling.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- 1. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of submission of tenders.**