

**HLL MOTHER AND CHILD CARE HOSPITALS LIMITED (HLLMCCH)
9/316, VIKAS NAGAR,
LUCKNOW, UTTAR PRADESH -226022**

Request for Proposals (RFP)

Scope of work: Supply of Hospital Furniture items for Twenty Mother & Child Care Hospitals in twenty different districts of Uttar Pradesh.

HLL Mother & Child Care Hospital Ltd., has entered into an agreement with Government of Uttar Pradesh for Operationalization of 100 beds Mother and Child Hospital in twenty (20) different districts of Uttar Pradesh under PPP model i.e. equip, finance, operate, maintain & transfer (EFOMT) basis.

HLL MCCH Ltd., invites sealed tenders in two bid systems (Technical Bid & Price Bid in two separate sealed envelopes) from manufacturers or authorized distributors/dealers for supply, assembling of Hospital Furniture & fixtures to twenty hospitals in twenty different districts of Uttar Pradesh as per specifications mentioned in Annexure –I of this document.

Tender Schedule

- All the bid needs to be submitted by **300 PM on Date: 5th February'2018** at the following address:

Team Leader- Administration & Commercial
HLL Mother & Child Care Hospitals Ltd.,
B-14A, 2nd Floor, Sector-62
Gautam Budh Nagar,
Noida-201307 (UP)

- **Pre-Bid Meeting** for any clarification or information would be held on **23rd January'2018** at above mentioned address.
- Technical Bids would be opened on **5th February'2018 at 4:00PM** at the same address.

Preparation of Tender

- The Manufacturer or Authorized distributors/dealers need to submit 'Technical Bid' & 'Financial Bid' separately in sealed envelopes by super scribing as "Proposal for Supply of Hospital Furniture to 20 Mother & Child Hospitals in different districts of Uttar Pradesh."

The "Techno – Commercial Tender" and "Price Tender" prepared by the bidder shall comprise the following:

A) Techno – Commercial Tender (Un priced Tender)

- Earnest money of Rs.10,00,000.00 (Rupees Ten Lakhs) in form on Demand Draft in favor of HLL Mother & Child Care Hospitals Ltd., payable at Noida.
- Distributors/dealers who quotes for goods manufactured by other manufacturer shall furnish Manufacturer's Authorization
- Power of Attorney/Authorization in favor of signatory of Tender documents.
- Documents and relevant details to establish that the goods and the allied services to be supplied by the bidder conform to the requirement of the tender.

- Confirmation that the supplied material shall be at par to the specification mentioned with respect to each item.
- Proof of certificate for turnover is required to be submitted with the technical bid (CA certificate or a copy of the Audited Balance Sheet, Profit & Loss account of immediately last preceding three years).
- Performance Statement along with relevant copies of purchase orders and end users' satisfaction certificate.
- Documentary evidence on Total Net worth of Manufacturer/Authorized Agent
- Certificate of Incorporation of the bidder.
- Affidavit that the firm is not banned by any Govt. body.
- List of Plant & Machinery
- Statement of deviations parameter wise from tendered technical specifications, if any.

B) Price Tender:

- The Bidder shall indicate on the Price Schedule all the specified components of prices shown therein including the unit prices and total tender prices of the goods and services it proposes to supply against the requirement.
- The rates quoted by the Bidder should be inclusive of all Taxes/Levies/Packaging & Forwarding, Freight [FOR at Districts mentioned in Annexure-II], assembling etc.
- The price quoted by the Bidder shall not be higher than the lowest price charged for the goods of the same nature, class or description to an individual/ firm/ organization or department of Govt. of India.

➤ The manufacturer or Authorized Agent shall bear all costs associated with the preparation and submission of bid.

Terms and conditions

1. The Bidder can be an individual/HUF/Association of Persons (AOP)/Society/Trust/Partnership firm/company incorporated under the laws of India. The required documents related to constitution/incorporation of entity like partnership deed/trusted/society by laws/memorandum and articles of association along with registration certificate of the entity should be closed.
2. No manufacturer or Authorized distributor/dealer shall be entitled to submit more than one bid whether jointly or separately. If one does so, all bids where in the manufacturer or Authorized distributor/dealer has participated shall stand disqualified.
3. The manufacturer or Authorized distributor/dealer should have an average annual relevant turnover of Rs. One Crore in the last three financial years to be able to qualify for bidding. Copies of latest ITCC, Annual Financial Report (Balance Sheet & Profit & Loss Account) for last three years to be enclosed.
4. The bidder must have PAN (Permanent Account Number) & GSTIN.
5. The bidder should provide detail profile of their work experience along with Xerox copies of Supply Orders /Purchase orders against which they successfully supplied similar furniture items to Govt. Hospitals/Private Hospital/Medical Institute etc.
6. The manufacturers or Authorized distributor/dealer applying would be required to submit Rs.10,00,000/- (Rupees Ten Lakhs) as Interest free Earnest Money Deposit (EMD) in the form of Demand Draft in favor of 'HLL Mother & Child Care Hospitals Ltd.' payable at Noida in the envelope containing Technical Bid. If the selected manufacturers or Authorized distributor/dealer are not able to supply the furniture as per delivery schedule, then the management reserves the right to forfeit the EMD. The EMD of the unsuccessful bidders would be returned in 45 days without any interest on EMD.
7. HLL MCCH Ltd., reserves rights to forfeit the EMD amount if at any stage any information/documents provided by the bidder are found forged/manipulated/modified.

8. HLL MCCH Ltd. reserves the right to accept /reject/select one or more than one bidder and to annul the bidding process any or all bids at any time prior to award of contract without thereby incurring any liability to the affected bidders.
9. The Bidder should have supplied in last 03 years from the date of Tender Opening, at least 75% of the quoted quantity of the similar furniture items meeting major specification parameters in India supported with documentary evidence.
10. HLL MCCH Ltd., reserves the right to assess the bidder capability and capacity to perform the contract satisfactorily before deciding on award of Contract.
11. HLL MCCH Ltd., reserves right to accept or reject any offer wholly or partly without assigning any reason thereof. The purchaser does not pledge itself to accept the lowest or any other offer and reserves to itself the right of acceptance of the whole or any part of the offer.
12. **Inspection:** HLL MCCH Ltd., reserves the right to inspect the factories of the bidders, at any point of time and also has the right to reject the tender/terminate/cancel the orders, based on adverse reports brought out during such inspections. The Bidder shall extend all facilities to the team to enable them to inspect the manufacturing process, quality control measures adopted etc. in manufacturing of items quoted/ordered.

HLL MCCH Ltd., reserves right to ask the technical qualified bidders to demonstrate products for inspection and final assessment of quality. In the event any products demonstrated by technically qualified bidder is found to be deficient in terms of quality, the order of such bidder shall be declared as technically not acceptable. Based on such assessment the final list of technically qualified bidders shall be drawn up.

13. Bidder qualifying the technical parameters would only be considered for opening of financial bids.
14. The basic rate quoted by the manufacturer or Authorized distributor/dealer should be valid for a period of one year from the date of award of purchase order as per details mentioned in Financial Bid document.
15. Bidders are expected to quote for all items or part thereof but not less than 50% of tendered items. Offers for items less than 50% of tendered items will be considered unresponsive and liable to be rejected.

16. The successful bidder shall not assign or transfer the contract to any other person or party. The tender is not transferable.
17. **Performance Security:** Within twenty-one (21) days from the date of the issue of notification of award by HLL MCCH Ltd., the supplier, shall furnish performance security for an amount equal to five percent (5%) of the total value of the bid quote in form of Account Payee Demand Draft or Bank Guarantee drawn/issued from any Scheduled Bank in India with validity up to one year after the date of completion of all contractual obligations by the supplier. HLL MCCH Ltd., will release the Performance Security without any interest to the supplier on supply & successful assembling of items. Bidder can furnish performance security after reducing the EMD submitted at the time of bidding.
18. In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the Purchase order without assigning any reason thereof and nothing will be payable by HLL MCCH Ltd., in that event the security deposit shall also stand forfeited.
19. HLL MCCH Ltd. reserves the right to award the purchase order to the L2 in the event L1 backs out after final discussions. In such case HLL MCCH Ltd., reserves the right to forfeit the EMD of the L1bidder. HLL MCCH Ltd., reserves the right to take any other suitable decision if required in order to complete the task as per project timeline.
20. The manufacturer or Authorized distributor/dealers needs to sign and stamp all pages of the technical bid along with the terms and conditions of tender.

21. Terms of Payment

On supply & assembling of furniture, State Level Committee will inspect & certify that the supplied furniture are as per specification given in tender. The payment to the supplier shall be released after 30 days of submission of report by the State Level Committee. However, in no case it would be delayed by 60 days from the date of supply & assembling of furniture items at particular facility on submission of following documents:-

- Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- Delivery Challan duly signed & stamped by the authorized representative of HLL MCCH Ltd;

- Bill/Invoice mentioning Permanent Account Number & GSTIN
- Documentary proof for supply & assembling of furniture items.
- Crossed/cancelled cheque of Bank Account for making payment through NEFT/RTGS.

22. Placement of Orders

HLL MCCH Ltd. Management reserves the rights:

- To place order to more than one manufacturer or Authorized distributor/dealer as per requirement.
- To place order for supply of Hospital furniture in phase manner for 20 Mother & Child Care Hospitals (or less in numbers). However, the tentative order placement schedule shall be as under:-

For Four Facilities	-	March'2018
For Six Facilities	-	June'2018
For Ten Facilities	-	September'2018

23. Quality & Specifications

Supplies by successful Bidder against Purchase Order shall be of the specified quality & specifications as incorporated at Annexure-I. The successful Bidder shall be fully responsible for the quality of the furniture items supplied. HLL MCCH Ltd shall be at liberty to claim compensation for any sub-standard/non-conforming supplies from the successful bidder and the bidder shall be liable to pay the same in addition to any other loss incurred. HLL MCCH Ltd., would also have the right to terminate the contract and/or forfeit the Earnest Money/Security Deposit without any further notice to the vendor and also to purchase the balance quantity at the entire risk and cost of the vendor.

HLL MCCH Ltd will not be responsible for any deterioration in the quality of the furniture items in transit.

The State Level designated Committee shall inspect to determine the supply of furniture items as per specification & standards and if it is reasonably anticipated or determined by the committee that the furniture items does not meets the specifications & standards, the supplier will be liable to remedy and rectify the defects or deficiencies or replace the furniture items, if required, at his cost. HLL MCCH Ltd reserves the right to forfeit the Earnest Money/Security

Deposit and also to purchase the said item at the entire risk and cost of the supplier, in case supplier denies remedying & rectifying the defects, deficiencies or replacement of furniture items.

24. Guarantee/Warranty

Bidder to confirm that the supplied furniture items will continue to conform to the description and quality for a period of 24 (Twenty Four) months or the manufacturer's/dealer standard guarantee/warranty whichever is more from the date of supply & installation of items at respective sites. If during the aforesaid guarantee/warranty period, any furniture item be found not giving satisfactory performance or has deteriorated, HLL MCCH Ltd. reserves the right to call upon the supplier to rectify the goods or such portion thereof as is found to be defective. In case supplier fails to rectify or replace the item within specified time, HLL MCCH Ltd. shall be entitled to recover the cost of that items with all expenses from the supplier.

25. Liquidated Damage/Termination

- HLL MCCH Ltd., reserves rights to deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods and/or services until actual delay or performance subject to a maximum of 5% of the contract prices. Once the maximum is reached, HLL MCCH Ltd may consider termination on the contract in case the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract/Purchase Order.
- In the event of termination of the contract in whole or in part, HLL MCCH Ltd., may procure goods and/or services similar to those cancelled, with such terms and conditions and in such manner as it deems fit and the supplier shall be liable to the HLL MCCH Ltd., for the extra expenditure, if any, incurred for arranging such procurement.

26. Force Majeure Clause

Delivery date will be extended without the vendor being subjected to penalty on account of liquidated damages only in the event of Force Majeure conditions within the contractual delivery period. Only the following will be considered as the causes of force majeure:

Act of GOD (Earthquake, Flood, Storm etc.), acts of States Direct & Indirect consequences of war (declared & undeclared), Hostilities, National Emergency, Civil Commotions & Strike (only those exceeding duration of 10 continuous days) or any other reason beyond the control of the vendor and accepted by HLL MCCH Ltd., The vendor shall immediately inform HLL MCCH Ltd., in writing supported by the documentary proof at the beginning and end of such impediments. It is to be understood that delivery schedule will be extended only for the duration of the above-mentioned impediments.

27. **Penalty for use of undue influence:** The Bidder should undertake that he will not give/offer/or promise to give directly or indirectly, any gift, consideration, reward, commission, fees, brokerage or inducement to any person in service of the HLL MCCH Ltd or otherwise in procuring the contract. Any breach of the aforesaid undertaking by the Bidder or any one employed by him shall entitle the Buyer to cancel the contract/reject the offer and recover from the Bidder the amount of any loss arising from such cancellation. Offering of any gift, bribe or inducement or any attempt at any such act on behalf of the bidder towards any officer/employee of the Buyer or to any other person in a position to influence any officer/employee of the Buyer for showing any favor in relation to this tender, shall render the Bidder to such liability/penalty as the Buyer may deem proper, including but not limited to termination of contract, imposition of penal damages, forfeiture of EMD.

28. **Arbitration :**

- All disputes or differences arising out of or in connection with the contract should be settled by bilateral discussions.
- Any dispute, disagreement or question arising out of or relating to this contract/relating to performance which cannot be settled amicably, to be referred by concern party to Chief Executive Officer, HLL Mother & Child Care Hospitals Ltd for appointment of Sole Arbitrator. The Arbitrator so appointed shall be a Government servant who had not dealt with matters to which this agreement relates and in course of his duties had not expressed views on all or any of the matter in disputes or differences. The Award of the Sole Arbitrator shall be final and binding on the parties.
- The venue of the arbitration shall be at Lucknow, Uttar Pradesh.

29. All the documents submitted should be self-certified & stamped by the manufacturers or authorized distributor/dealer.

30. The Bidder should also submit an undertaking duly signed & Stamped as per Annexure-III

We agree and abide by all terms and conditions as mentioned above including the validity of the offer.

Utmost confidentiality of the data provided shall be maintained.

HLL Mother & Child Care Hospitals Limited
9/316, Vikas Nagar,
Lucknow, Uttar Pradesh-226022

Annexure-I: Technical Bid Format

	SPECIFICATION(TECHNICAL)	MANUFACTURER OR AUTHORIZED DISTRIBUTOR/DEALER REPOSE (Agree or Disagree)
1.	Name of the firm/Society/ Company/Proprietary Concern	
2.	Address of registered office(along with telephone number, if any)	
3.	Address of the office at Delhi/NCR (if any) (along with telephone number, if any)	
4.	Earnest Money Deposits (EMD) submitted	Yes Or No
5.	PAN No.	
6.	GSTIN No.	
7.	Manufacturers or Authorized Distributor/Dealer agree to provide services allied and incidental to the equipments supplied.	Yes Or NO
8.	HLLMCCH's authorized representatives may undertake activities such as measuring, examining, testing, gauzing one or more characteristics of the equipments supplied.	

S.No.	Item	Specifications
Hospital Furniture		
<u>Technical Specification of Fowler Bed with Mattress Drip/Saline Stand</u>		
1.	Hospital Bed	<p><u>Fowler Bed</u></p> <p>1. DIMENSION:</p> <ol style="list-style-type: none"> Length (Overall) = 2015mm +/- 5 mm variation Width (Overall) = 965 mm +/- 5 mm variation Bed Top Height min 490 mm and max 710mm (without mattress) <p>2. BED CONSTRUCTION:</p> <ol style="list-style-type: none"> Bed should consist of sections namely: Bed Frame, Head Section, fixed section. Bed frame is made of rectangular pipe of size 60 mm x 30mm x 1.6mm thick. Trendelenburg / Reverse Trendelenburg by Crank System. Head section is capable of being raised to the angle of 70 degree from the horizontal. The head section slope adjustment is achieved by separate screw mechanism with thrust bearing and proper lubrication. The rest of bed surface is fixed. Head section is made of CRCA 1.25mm thick uniformly perforated Sheet welded with the outer tube 1.6mm thick x 25.4mm dia. The rest of the bed surface which is fixed is also uniformly perforated and welded to the rectangular frame.
		<ol style="list-style-type: none"> Provision of saline stand on both sides of Beds. Should be supplied with mattress (sectional). <p>3. Nuts & Bolts:</p> <ol style="list-style-type: none"> All Fasteners like nut and bolts etc are electro-galvanized. <p>4. Head & Foot Boards:</p> <ol style="list-style-type: none"> Head and Foot boards are made of 31.75 mm dia x 1.25 mm thick stainless steel tube fitted with laminated panel board with help of stainless steel bracket. The thickness of laminated panel board not less than 18 mm (+/- 1mm). Both the head and foot boards shall be detachable for easy access to patients. The overall height of head and foot board shall be 250 mm on both sides. <p>5. Finishing:</p> <ol style="list-style-type: none"> Steel welding wherever required, is being done by MIG welding process to minimize distortion and for deep penetration of the weld. All the steel components are pre-treated for degreasing, de-rusting and phosphating. After proper pre-treatment, the steel components are epoxy powder coated and over baked at temp. Above 200 deg. C to provide scratch resistance surface coating film of thickness 45-50 microns. <p>6. Side Rails:</p> <ol style="list-style-type: none"> Drop side railing provided on both sides shall be made of 19.0 mm dia x 1.2 mm thick CRC pipe. The length of side railing shall be 122 cm & width 38 cm with 4 equal vertical components. The sliding side shall have up/down movement & shall be fixed on SS 304. <p>7. Wheels:</p> <ol style="list-style-type: none"> The bed shall be mounted on four number SS fork castors of 100 mm x 32 mm & shall be made out of semi round rubber hardness 85/90, Pressed fitted over the virgin polyamide pulling. The S.S material asked for in any specification should be of S.S 304 Grade.

			<p>8. Standards & Safety :</p> <p>a) Should be BIS or US FDA or EU (CE) approved.</p> <p>b) Manufacturer should have ISO 13485 certificate for quality standard.</p> <p>c) Any warning signs would be adequately displayed.</p>
2.	ICU Bed		<p>ICU Bed with IV stand</p> <ul style="list-style-type: none"> • Bed should be Four-crank with tilting. The Bed should include two pieces integral Spit safety side rails a Standard Mattress, IV Rod.
			<ul style="list-style-type: none"> • The ICU Bed Should be specially designed for Paediatrics • Length: 2100 mm to 2250mm; Width: 950 mm to 1000 mm; Height : 400 to 750mm • Backrest Tilting: $60^{\circ} \pm 5^{\circ}$ AND Knee Board tilting : $35^{\circ} \pm 5^{\circ}$ Reverse and Trendelenburg $12^{\circ} \pm 2^{\circ}$; • The bed should be painted with anti-rust paint. Flexible bedstead built in crank controlled. • Crank features universal linking cardio structure, safe, quiet, smooth operation. Built in design crank handle when at lowest position, Four bed legs should provide enable stable position of the bed. • The bedside End plate should be removable and lockable. • The metallic hanging objects tracks, supporting all kinds of hooks and fixed drainage bags, urinals and other objects can be placed, and can move freely position. • Bed's structural parts should have anti bacteria double powder coating treatment or should be easy to clean for better infection control. • Bed should have two pieces individual side rails. • Integral forming head/foot end board, available for CPR. • Bed should have quiet and smooth while operating and have easy maintenance and also should have good quality castors. • Safe working Load should be 100 Kg. • Demonstration of the system is essential. • Manufacturer or supplier should have ISO13485. • Should be USFDA or EU (CE) approved.
3.	Eclampsia bed		<ol style="list-style-type: none"> 1. Overall Size 2180mm L x 1000mm W x 600mm H +-5% 2. The main frame should be made from 60mm x 30mm x 16 G ERW rectangular tubes. 3. Four sections top should be made from uniformly perforated 18 G C.R.C.A sheets and should be suitably fitted to the main frame 4. Back rest and knee rest should be maneuvered by the screw mechanism with one common handle 5. Detachable head and leg bows: should be made from stainless steel 6. 31.7mm dia x 18 G tube with laminated panel of approx. size 810mm L x 140mm wide x 14mm thick on stainless steel bracket. the panels should be cushioned for safety 7. Should be Eu(CE) or BIS approved. 8. Manufacturer or supplier should have ISO-13485 certificate.

4.	Bed side locker		<ul style="list-style-type: none"> • Overall Dimension: L 410 mm x W 410 mm x H 820 mm • Locker are made from machine pressed CRCA steel sheets enclosed on three sides having one drawer on top, one open shelf and one cabinet belong with swing door, Stainless steel top with edges raised on three sides is affixed on top of the locker with high class bonding agent. The whole body is mounted on four swivel ball castors of 50 mm dia. • Buffers are provided at rear of lockers to prevent the damage to ward wall. • Drawer inside size: 310 mm (W) x 360 mm (D) x 100 mm (H) ±10% • Open shelf closed on 3 sides: 390mm (W) x 390 mm (D) x 175 mm (H) ±10% • Lower cabinet size: 390 mm (W) x 375 mm (H) x 390 mm (D) ±10% • Steel welding wherever required, is done by MIG welding process to minimize distortion and for the deep penetration of the weld. • All the steel components are pre-treated for de-greasing, de-rusting and Phosphating. After proper pre-treatment, the steel components are epoxy powder coated and oven baked at temp. Above 200 deg. C to provide scratch resistance surface coating of film thickness 45-50 microns. • The S.S material asked for in this specification should be of S.S 304 Grade of 22G in matt finish. • Should be BIS approved. • Manufacturer should have ISO 13485 certificate for quality standard.
5.	Bed side table		<ul style="list-style-type: none"> • Over Bed Table Pre laminated Top Operate Crank • Wooden top adjustable by gear • All metal elements are pre treated and epoxy powder coated • Dimension: 760mm L X 400 mm W X 760-1050 mm H • European CE certification or USFDA certification or BIS certification • Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality
6.	Bed side stool for attendant		<ul style="list-style-type: none"> • Square tubular framework. • Pre-treated Epoxy powder coated finish. • S.S. Top: (30 x 30) cms. Ht. 45cm. ±5% • European CE certification or USFDA certification or BIS certification

S.No.	Item		Specifications
			<ul style="list-style-type: none"> Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality
7.	Attendant bench		<ul style="list-style-type: none"> Overall size should be 180Lx60Wx46H cm \pm 5% Framework should be made up of rectangular MS tubes. The fabric should be pressed artificial leather Legs fitted with PVC stumps Finish should be pre treated& epoxy powder coating European CE certification or USFDA certification or BIS certification Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality
8.	Bed side screen		<ul style="list-style-type: none"> Overall size should be 1680x2450 mm. Should be made of MS tubular frame work, 19mmx18G in three section Legs made of 25mm x18G mounted on 5cms castors with hooks & springs along with curtains Each span should be 810mm wide. The finish should be Pre-treated and epoxy powder coated. European CE certification or USFDA certification or BIS certification Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality
9.	Wash basin stand		<ul style="list-style-type: none"> Complete Stainless Steel Construction Stainless Steel Basin Bowl - 360mm Diameter x 100mm Deep Castors : 3 x 75mm Grey Rubber Height : 100mm Tubular Steel Framework. Finishing : Epoxy Powder Coated Quality European CE certification or USFDA certification or BIS certification. Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality
10.	Instrument Almirah		<ul style="list-style-type: none"> Made of CRCA sheets provided with single doors and glass shelves Pre treated and Epoxy powder coated

S.No.	Item		Specifications
			<ul style="list-style-type: none"> • Dimension: 63" H X 24" W X 18"D with 5 glass shelves
11.	Bed pans		<ul style="list-style-type: none"> • Sterilisable / steel make standard. • Size :Pan Size in mm - 375 x 285 x 97±5% • Female - Perfection type, Seamless
12.	Medicine Almirah		<ul style="list-style-type: none"> • Made of CRCA sheets provided with single doors and glass shelves • Provided with two polystyrene boxes each with three lockable drawers • Pre treated and Epoxy powder coated • Provided with three CRCA sheets / Glass shelves • Provided with SS IV rods and corner buffers and cardiac massage board • The cart is mounted on 125 mm dia castors with two brakes • Dimension: 63" H X 24" W X 18"D with 5 glass shelves
13.	Dressing trolley		<ul style="list-style-type: none"> • Overall approximate dimension: 1010 mm L X 510 mm W X 900 mm H. • Approximate shelf dimension: 750 mm L X 500 mm W. • S.S. tubular frame mounted on four 125 mm diameter castors with synthetic body, two with brake & two without brake. • Two S.S. shelves with protective railings on all four sides. • With S.S. bowl and S.S. bucket. • Supplied in SKD condition. • Should be EU CE or BIS approved. • Manufacturer should have ISO 13485 certificate for quality standard.
14.	Medicine trolley		<ul style="list-style-type: none"> • Frame work made from SS steel material. All stainless steel should be 304 grade/16 G • Flat top of SS and at least 6 inch deep removable bucket at bottom • Should have multiple long drawers to hold drug strips made of high quality epoxy plastic or steel material with convenient and smooth slide in and slide out motion (At least 28-32 separate drawers in about six to eight row) • The front of each drawer should be half covered on which removable medicine label can be pasted and upper half open to see content inside • Approx. size 750(L) x 450(W) x 850(H) mm±10% • Should be fitted with good quality swivel noiseless castor of 100mm dia fitted with brakes • Equipped with waste bin, needle disposable container, file cassette & guard rails

S.No.	Item		Specifications
			<ul style="list-style-type: none"> • Equipped with lock key system • Swivel noiseless castors with brakes 10 cm diameter • Should be Eu(CE) or BIS approved. • Manufacturer or supplier should have ISO-13485 certificate.
15.	Equipment trolley		<ul style="list-style-type: none"> • Frame work constructed with heavy duty MS Steel sheet & Angle • Grade of steel should be SS 304 • Overall size approximate 1800 L x 1050 W x 450 D (mm) ±10% • Two glass doors with individual latch handles and key locking facility • Should be Eu(CE) or BIS approved. • Manufacturer or supplier should have ISO-13485 certificate.
16.	Food service trolley		<ul style="list-style-type: none"> • Dimension: 32" x 28" x 42" ±5% • The trolley should be mounted on 100 mm dia castors • European CE certification or USFDA certification or BIS certification • Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality
17.	Linen trolley		<ul style="list-style-type: none"> • Overall size: 910mm H x 510mm dia. SS tubular framework fitted with three swivel castors, 100mm dia • Framework made of 25.4mm dia x 18 G verticals, upper ring made of 19mm dia x 18 G tubes and support stays of 19 mm dia. x 18 G stainless steel tubes • Should be supplied with canvas bag • All SS components should be of 304 quality • Should be Eu(CE) or BIS approved. • Manufacturer or supplier should have ISO-13485 certificate.
18.	Streture with trolley		<ul style="list-style-type: none"> • Overall approximate dimension: 2030 mm x 560 mm x 820 mm, 80" x 22" x 32" (L x W x H) • S.S. tubular framework mounted on four swivel castor of 200 mm diameter, two with brake and two without brake • Removable stretcher top made of stainless steel sheet supported on S.S. tubular frame • Two provisions for IV. Pole

S.No.	Item		Specifications
			<ul style="list-style-type: none"> • Supplied with safety belts • Safe working load of 135 kgs and patient load bearing capacity of 130 kgs • Rexine covered 25 mm thick single section mattress 1830 mm x 560 mm, 72" x 22" (L x W) • With I.V. rod • S.S. oxygen cylinder cage (0166G) • S.S. storage tray (0149A) • Manufacturer should have ISO 13485 certificate for quality standard. • Should be Eu(CE) or BIS approved. • Manufacturer or supplier should have ISO-13485 certificate.
19.	Wheel chair		<ul style="list-style-type: none"> • Overall size: 68Wx 108Dx 98H ±10% • Detachable armrest & footrest. • Folding padded seat & back • Rear Tyres with break and front castors of 20cm dia • Supplied with common arrangements • European CE certification or USFDA certification or BIS certification • Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality
20.	Oxygen trolley		<ul style="list-style-type: none"> • Cylinder trolley (Push type) fitted with 2 castors of 100mm dia with MS body frame • Suitable for B type oxygen gas cylinders • Trolley with SS base • Should not wobble on flat ground when placed straight • Should be pre-treated chemically and metal surface should then be coated with epoxy polyester powder with paint film thickness of 60 microns (minimum) and over baked at 180-200° C • Should be Eu(CE) or BIS approved. • Manufacturer or supplier should have ISO-13485 certificate.
21.	IV stand		<ul style="list-style-type: none"> • CRCA pipe mounted, pre treated epoxy powder coated and mounted on 100 mm castors
22.	Emergency drug Cart		<u>Technical Specification of Emergency Drug Cabinet Trolley</u>

S.No.	Item		Specifications
			<ul style="list-style-type: none"> • It should be ABS body, with aluminium columns. • Should be fitted with needle disposal holder, IV Pole, • Following should be provided in the trolley: <ul style="list-style-type: none"> ○ Dust Basket ○ Utility container ○ File bag ○ Defibrillator shelf ○ Power outlet & Hooks ○ Oxygen tank holder ○ CPR board • The trolley should be with sliding side shelf, SS guard rail, conclave top covered transparent soft plastic glass and centralised lock. • It should provide with 2 small & 3 big drawers with partitions. • ABS basket & drawer frame should be well repositioned easily. • Good quality casters, two with brakes to be fitted. • The size should be approx. 850x520x950mm • European CE certification or USFDA certification or BIS certification • Quality Certificate of manufacturer like OHSAS-18001 or ISO 9001- 14001/9001-2008/ 9001-13485 should be attached to ensure quality • IV Rods brackets, Corners to be buffered, Castors 100mm dia (NOT 50MM) with 4 Brakes (NOT 2) , 3 Lockable Drawers, 6x ABS Baskets multi colored removable bins."
23.	Staff chairs for OPD and wards		<ul style="list-style-type: none"> • Standard wheel / fixed chair as per colour code of hospital
24.	Nursing station/ OPD /ward table		<ul style="list-style-type: none"> • Standard with Crash Cart trolley, Linen Almirah, Medicine Almirah, Computer, printer and other standard printing stationeries, chair, table etc.
25.	Visitor chairs (5 seater) for waiting area		<ul style="list-style-type: none"> • Made up of perforated sheet, Chrome plated finish for soothing look • Dimension 69mm L X 25 mm W X 30 mm H

List of Hospitals/Districts	
1	Ambedkar Nagar
2	Auriya
3	Azamgarh
4	Bagpat
5	SantKabir Nagar
6	Etah
7	Hardoi
8	JyotibaPhule Nagar (Amroha)
9	Kannauj
10	Kaushmabi
11	Kushinagar
12	Maharajganj
13	Siddharth Nagar
14	Mau
15	Pratapgarh
16	Jaunpur
17	Gorakhpur
18	Ghazipur
19	Deoria
20	Ballia

Date :

Place :

Signature of Authorized Person

Full Name:

Company's Seal:

UNDERTAKING

This has reference to the tender published in the website of HLL Mother & Child Care Hospitals Ltd., on..... In response to the RFP, we have submitted our technical & financial bids on.....at your office..... In connection with the above bids, we hereby declare as under:-

1. That we are neither related to any of your Trustees, Officers and other employees nor do we have any financial, commercial or other interests with any of the above persons in any capacity whatsoever.
2. That we have submitted the bids in the name of M/s.....and declare that no other bids have been submitted by us in the name of any other firms/companies/proprietors/individuals which comes under the same management and related parties.
3. We hereby undertake that the information, documentary evidence on financial standing, past experience, performance report from clientele provided with tender are authentic. HLL MCCH Ltd reserves the sole right to forfeit the EMD submitted by us if at any stage the information, documents provided by us, are found forged/manipulated/modified.
4. We here by undertake that in case of any violations to the above declarations at any stage of the contract, HLL MCCH Ltd. reserves the sole right to cancel the contract and recover the full value of the contract from us.

Signature of Authorized Person

Date :

Full Name:

Place :

Company's Seal

